

Application Notes

NDIR - Determination of Linearization Coefficients for Gas Analysis Using KEMET Dual Sensors



Table of Contents

1	INTRODUCTION	.2
2	CALIBRATION DATA	.3
3	USING EXCEL TO DETERMINE LINEARIZATION COEFFICIENTS	.5
4	MODIFIED BEER LAMBERT LAW	.9

DOCUMENT HISTORY

Version Date Ch		Change Ref.	Change Details
01 8 JUN 2020		N/A	First Release

1 INTRODUCTION

This document explains the method for setting the linearization coefficients for calculating the concentration of a gas being analysed. This method requires a calibration procedure of introducing known concentrations of the gas being analysed into the sample cell and recording the signal strengths for each concentration.

This document should be used alongside the excel spreadsheet "KEMET NDIR Gas Coefficients".



2 CALIBRATION DATA

Using a dual sensor set up involving a gas sensor and a reference sensor the following data is recorded by taking the root mean squared values for each sensor's output.

	Measured				
		Sensor data			Т
	Gas Concentration %	Ref	Gas	Ratio (Gas/Ref)	Transmittance
sample1	0	21903	27976	1.277	1.00000
sample2	10	21931	27519	1.255	0.98241
sample3	20	21929	27065	1.234	0.96629
sample4	30	21950	26681	1.216	0.95167
sample5	40	21933	26322	1.200	0.93959
sample6	50	21945	25987	1.184	0.92713
sample7	60	21961	25744	1.172	0.91779
sample8	70	21971	25546	1.163	0.91031
sample9	80	21978	25347	1.153	0.90293
sample10	90	21982	25205	1.147	0.89771
sample11	100	21971	25068	1.141	0.89328

Table 1 – Calibration Data

Using the definition of transmittance

$$T = \frac{I}{I_o}$$

Where 'l' is the measured signal and 'l_o' is the signal with 100% transmission (zero signal absorbed by the gas). Dividing the gas by the reference normalizes the output for the system being calibrated, with the value of gas/ref with 0% concentration of gas giving the value of 'l_o' for the system since by definition there is 100% transmission when there is no gas present to absorb the IR.

Then the amount of IR absorbed is then the total amount of IR emitted (I_0) minus the amount transmitted. This is normalized to a ratio by dividing both terms by I_0 .

$$A = \frac{I_o - I}{I_o} = 1 - \frac{I}{I_o}$$

This definition of absorption is taken from "Non-Dispersive Infrared Gas Measurement" by J. Y. Wong and R. L. Anderson. It is really the combination of signal losses from photon absorption and scattering.



Т	1-T
Transmittance	Absorption
1.00000	0.00000
0.98241	0.01759
0.96629	0.03371
0.95167	0.04833
0.93959	0.06041
0.92713	0.07287
0.91779	0.08221
0.91031	0.08969
0.90293	0.09707
0.89771	0.10229
0.89328	0.10672

Table 2 – Transmittance and Absorption

Plotting the absorption against the concentration from calibration gases gives the following graph.



Figure 1 – Absorption vs. Concentration

The above graph is an exponential relationship however each system will be different so a few coefficients for calibration are required. The following equation for absorption is used.

$$ABS = a(1 - e^{-bx^c})$$

This is the format of the equation that the values of a, b and c will be determined from. This is due to the fact that the gas concentrations are the inputs for calibration process and the reduction in signal strength is the output of the calibration procedure.

The above information is believed to be correct but does not purport to be all inclusive and must be used only as a guide. © KEMET Electronics Corporation • KEMET Tower • One East Broward Boulevard

NDIR - Determination of Linearization Coefficients for





3 USING EXCEL TO DETERMINE LINEARIZATION COEFFICIENTS

In order to determine the coefficients, the add-in of excel 'Solver' must be enabled. This is done as shown below.









3. Click 'Go...' next to the manage Excel add-ins.

M <u>a</u> nage:	Excel Add-ins	\mathbf{v}	<u>G</u> o

4. Ensure that the 'Solver' is selected in the Add-Ins available box. Then click 'OK'.

Add-Ins	?	×		
<u>A</u> dd-Ins available:				
Analysis ToolPak	OK	(
■ Euro Currency Tools	Cano	cel		
9mmi	Brows	se		
	A <u>u</u> toma	tion		
~ ·				
Solver Add-in				
Tool for optimization and equation solving				

The solver options will now be available in the Analysis section of the top bar of Excel.



This is the tool that will be used to iterate towards values of the coefficients to provide a best fit curve. There are additional steps to set-up the spreadsheet so that the solver tool has what it needs to converge on an answer.

The solver tool needs to know the calculated values of absorbance and the actual values from calibration. Using these values as two columns create a third column that will be the square of the difference between the two columns.

The above information is believed to be correct but does not purport to be all inclusive and must be used only as a guide.

 $\textcircled{\sc order {\sc order of the sector}}$ $\sc order \r{\sc order order order}$ $\sc order \r{\sc order order}$ $\sc order order order$ $\sc order$



	Predicted	
1-T	Predicted	
Absorption	Absorption	diff^2
0.00000	0.0000	
0.01759	1.0000	9.65E-01
0.03371	1.0000	9.34E-01
0.04833	1.0000	9.06E-01
0.06041	1.0000	8.83E-01
0.07287	1.0000	8.60E-01
0.08221	1.0000	8.42E-01
0.08969	1.0000	8.29E-01
0.09707	1.0000	8.15E-01
0.10229	1.0000	8.06E-01
0.10672	1.0000	7.98E-01

Table 3 – Layout of Actual Absorbance and Predicted Absorbance

Solver' parameters

8.64E+00	<- 'Set objective'	(1)
8.64E+00	<- 'Set objective'	(1)

Finally add a final box that will contain the summation of the squares of the differences. This summation of squared errors is used by the solver tool.

The final elements required are the boxes containing a, b and c.

Table 4 – Cells for Coefficients

Coefficients	value	
а	1	<- 'by changing
b	0	variable cells' - select
С	1	ali three (2)

The predicted column shown in table x is then assigned the equation

$$ABS = a(1 - e^{-bx^c})$$

with the concentration values coming from the calibration process, and the values for a, b and c coming from boxes created to contain them.

Starting the values for a and c at 1 and b set to zero, this will allow the solver to converge on an answer. These initial values are an example and may not be the best for every system. However, they work well the system used for this calibration procedure.



Click solver on the top	bar of excel to o	pen the solver o	ptions screen.
-------------------------	-------------------	------------------	----------------

		Solver Paramet	ers	
Se <u>t</u> Objective:				E
To: <u>M</u> ax	◉ Mi <u>n</u>	○ <u>V</u> alue Of:	0.99	
By Changing Variable Ce	ells:			
				1
Subject to the Constrain	ts:			
			^	<u>A</u> dd
				<u>C</u> hange
				<u>D</u> elete
				<u>R</u> eset All
			~	Load/Save
Make Unconstrained	Variables No	on-Negative		
Select a Solving Method:	GF	G Nonlinear	¥	O <u>p</u> tions
Solving Method				
Select the GRG Nonline Simplex engine for line problems that are non-	ar engine fo ar Solver Pro smooth.	r Solver Problems that blems, and select the	at are smooth nonlir e Evolutionary engin	near. Select the LP e for Solver
<u>H</u> elp			<u>S</u> olve	Cl <u>o</u> se

We want to minimise the summed error squared value. This is termed the 'Objective'. Set the objective to the cell containing the squared error summation. And ensure that the selection of minimum is checked.

Next we need to set the cells to be altered to help converge on a particular set of coefficients for the minimum error. This is the box with 'By Changing Variable Cells' above it. Select the three cells that contain the coefficients for this box.

Now click 'Solve'. This will do one iteration of the algorithm. Repeat the above procedure until the values converge.

The values of coefficients determined from the above procedure are shown below.

NDIR - Determination of Linearization Coefficients for

Gas Analysis Using KEMET Dual Sensors

Table 5 – Determined Values of Coefficients

	Predicted	
1-T	Predicted	
Absorption	Absorption	diff^2
0.00000	0.0000	
0.01759	0.0171	2.33E-07
0.03371	0.0337	5.55E-11
0.04833	0.0485	2.15E-08
0.06041	0.0613	8.70E-07
0.07287	0.0724	2.20E-07
0.08221	0.0818	1.36E-07
0.08969	0.0899	2.70E-08
0.09707	0.0966	2.07E-07
0.10229	0.1023	6.88E-11
0.10672	0.1071	1.17E-07

Solver' parameters

1.83E-06 <- 'S	Set objective' (1)
----------------	--------------------

Coefficients	value	
а	0.13024248	<- 'by changing
b	0.01148979	variable cells' - select
С	1.08837583	all three (2)

4 MODIFIED BEER LAMBERT LAW

The above procedure has been based around the equation

$$ABS = a \big(1 - e^{-bx^c} \big)$$

however, this is not an equation that outputs concentration values. Rearrange this equation to solve for 'x'. Which is the concentration.

$$x = \sqrt[c]{\frac{\ln\left(1 - \frac{ABS}{a}\right)}{-b}}$$

This is now the equation for determining the concentration of gas in the cell. With the values of a, b and c having been determined through the calibration process and using excel solver.

Ĵ

 $[\]ensuremath{\textcircled{\sc b}}$ KEMET Electronics Corporation ${\ensuremath{\cdot}}$ KEMET Tower ${\ensuremath{\cdot}}$ One East Broward Boulevard



NDIR – Determination of Linearization Coefficients for Gas Analysis Using KEMET Dual Sensors





Measured Gas Concentration %	Predicted Gas Concentration %	R^2
0	0.000	0.99979
10	10.279	
20	20.005	
30	29.894	
40	39.223	
50	50.461	
60	60.426	
70	69.777	
80	80.739	
90	89.984	
100	99.219	

Table 6 - Table of Actual Concentration vs. Predicted

This process produces an accurate equation for predicted values of gas concentration.